

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2592</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>PARSONS Engineering of New York, I</u>	Datafile :	<u>VX047022.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0716WBS01	Vinyl chloride	20	16.6	ug/L	83			65	117	
	1,1-Dichloroethene	20	17.7	ug/L	89			74	110	
	2-Butanone	100	140	ug/L	140		*	65	122	
	Carbon Tetrachloride	20	18.4	ug/L	92			77	113	
	Chloroform	20	20.0	ug/L	100			79	113	
	Benzene	20	18.5	ug/L	93			82	109	
	1,2-Dichloroethane	20	20.1	ug/L	101			80	115	
	Trichloroethene	20	17.6	ug/L	88			77	113	
	Tetrachloroethene	20	17.5	ug/L	88			67	123	
	Chlorobenzene	20	19.0	ug/L	95			82	109	