

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: PARS02
 Lab Code: ACE SDG NO.: Q2592
 Lab File ID: VX047016.D Date Analyzed: 07/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:26
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	329194	5.56	557691	6.76	489553	10.06
UPPER LIMIT	658388	6.056	1115380	7.263	979106	10.555
LOWER LIMIT	164597	5.056	278846	6.263	244777	9.555
EPA SAMPLE NO.						
WC-SOIL-20250711	432741	5.56	774714	6.77	742893	10.06
VX0716WBL01	324808	5.56	553162	6.77	506524	10.06
VX0716WBS01	322183	5.57	547259	6.77	491570	10.06
VX0716WBSD01	310963	5.56	532754	6.77	475577	10.06

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.