

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>PARS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2592</u>
Instrument ID:	<u>MSVOA_W</u>	Calibration Date/Time:	<u>07/14/2025 09:17</u>
Lab File ID:	<u>VW031836.D</u>	Init. Calib. Date(s):	<u>06/30/2025 06/30/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>09:54 12:55</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.341	0.280		-17.89	20
Chloromethane	0.411	0.378	0.1	-8.03	20
Vinyl Chloride	0.540	0.531		-1.67	20
Bromomethane	0.421	0.400		-4.99	20
Chloroethane	0.367	0.358		-2.45	20
Trichlorofluoromethane	0.490	0.456		-6.94	20
1,1,2-Trichlorotrifluoroethane	0.544	0.533		-2.02	20
1,1-Dichloroethene	0.596	0.589		-1.17	20
Acetone	0.177	0.145		-18.08	20
Carbon Disulfide	1.594	1.487		-6.71	20
Methyl tert-butyl Ether	1.011	1.128		11.57	20
Methyl Acetate	0.507	0.534		5.32	20
Methylene Chloride	0.814	0.750		-7.86	20
trans-1,2-Dichloroethene	0.633	0.635		0.32	20
1,1-Dichloroethane	1.156	1.213	0.1	4.93	20
Cyclohexane	1.022	0.980		-4.11	20
2-Butanone	0.231	0.228		-1.3	20
Carbon Tetrachloride	0.481	0.486		1.04	20
cis-1,2-Dichloroethene	0.728	0.791		8.65	20
Bromochloromethane	0.510	0.541		6.08	20
Chloroform	1.228	1.308		6.51	20
1,1,1-Trichloroethane	0.946	0.958		1.27	20
Methylcyclohexane	0.587	0.596		1.53	20
Benzene	1.397	1.493		6.87	20
1,2-Dichloroethane	0.472	0.504		6.78	20
Trichloroethene	0.349	0.356		2.01	20
1,2-Dichloropropane	0.338	0.370		9.47	20
Bromodichloromethane	0.511	0.557		9	20
4-Methyl-2-Pentanone	0.295	0.307		4.07	20
Toluene	0.894	0.965		7.94	20
t-1,3-Dichloropropene	0.480	0.537		11.88	20
cis-1,3-Dichloropropene	0.544	0.609		11.95	20
1,1,2-Trichloroethane	0.286	0.308		7.69	20
2-Hexanone	0.200	0.211		5.5	20
Dibromochloromethane	0.341	0.372		9.09	20
1,2-Dibromoethane	0.286	0.299		4.55	20
Tetrachloroethene	0.327	0.316		-3.36	20
Chlorobenzene	1.103	1.152	0.3	4.44	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>PARS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2592</u>
Instrument ID:	<u>MSVOA_W</u>	Calibration Date/Time:	<u>07/14/2025</u> <u>09:17</u>
Lab File ID:	<u>VW031836.D</u>	Init. Calib. Date(s):	<u>06/30/2025</u> <u>06/30/2025</u>
Heated Purge: (Y/N)	<u>Y</u>	Init. Calib. Time(s):	<u>09:54</u> <u>12:55</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.913	2.034		6.32	20
m/p-Xylenes	0.735	0.786		6.94	20
o-Xylene	0.681	0.751		10.28	20
Styrene	1.177	1.279		8.67	20
Bromoform	0.210	0.216	0.1	2.86	20
Isopropylbenzene	3.803	4.171		9.68	20
1,1,2,2-Tetrachloroethane	0.844	0.921	0.3	9.12	20
1,3-Dichlorobenzene	1.701	1.869		9.88	20
1,4-Dichlorobenzene	1.741	1.813		4.14	20
1,2-Dichlorobenzene	1.555	1.670		7.39	20
1,2-Dibromo-3-Chloropropane	0.165	0.166		0.61	20
1,2,4-Trichlorobenzene	0.959	0.982		2.4	20
1,2,3-Trichlorobenzene	0.881	0.953		8.17	20
1,2-Dichloroethane-d4	0.718	0.717		-0.14	20
Dibromofluoromethane	0.326	0.332		1.84	20
Toluene-d8	1.214	1.228		1.15	20
4-Bromofluorobenzene	0.447	0.466		4.25	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.