

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: PARS02
 Lab Code: ACE SDG No.: Q2592
 Instrument ID: MSVOA_W Calibration Date/Time: 07/14/2025 09:17
 Lab File ID: VW031836.D Init. Calib. Date(s): 06/30/2025 06/30/2025
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:54 12:55
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	53.40	-6.807999	20