

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: PARS02
Lab Code: ACE SDG NO.: Q2592
Lab File ID: VW031836.D Date Analyzed: 07/14/2025
Instrument ID: MSVOA_W Time Analyzed: 09:17
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	228359	7.96	408363	8.85	364073	11.63
UPPER LIMIT	456718	8.459	816726	9.349	728146	12.129
LOWER LIMIT	114180	7.459	204182	8.349	182037	11.129
EPA SAMPLE NO.						
WC-SOIL-20250711	205400	7.97	386707	8.86	368231	11.64
VW0714SBL01	179274	7.96	359991	8.85	334968	11.64
VW0714SBS01	217952	7.96	419177	8.85	378051	11.63
VW0714SBSD01	213196	7.97	419599	8.86	383633	11.64

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.