

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: PARS02
 Lab Code: ACE SDG NO.: Q2592
 Lab File ID: VW031836.D Date Analyzed: 07/14/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:17
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	165994	13.556				
UPPER LIMIT	331988	14.056				
LOWER LIMIT	82997	13.056				
EPA SAMPLE NO.						
WC-SOIL-20250711	164076	13.56				
VW0714SBL01	154203	13.56				
VW0714SBS01	171979	13.56				
VW0714SBSD01	180632	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.