

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH QC BATCHID: LB136487

	NA	NO	YES
1.Calibration Summary met criteria.	_____	_____	_____✓
2.ICP Interference Check Sample Results Summary Submitted.	_____	_____	_____✓
3. Serial Dilution Summary (if applicable) Submitted.	_____	_____	_____✓
4.Laboratory Control Sample Summary (if applicable) Submitted.:	_____	_____	_____✓
5.Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓	_____
6.Matrix Spike/Matrix Spike Duplicate Recoveries Met Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓
7.Sample Duplicate Analysis Met QC Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓
8.Digestion Holding Time Met If not met, list number of days exceeded for each sample:	_____	_____	_____✓
9.Analysis Holding Time Met If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓

ADDITIONAL COMMENTS:

1) Q2485-01 : MS/MSD fail for more than 50% parameters due to unknown interferences during digestion with sample matrix.

Analyst jaswal

Date 7/16/2025 12:00:00 AM

QA REVIEW

Date