

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client:	<u>Tully Environmental, Inc</u>	SDG No.:	<u>Q2593</u>
Contract:	<u>TULL01</u>	Lab Code:	<u>ACE</u>
ICS Source:	<u>EPA</u>	Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	99600	101000	99	85600	116500	07/15/2025	13:29	LB136487
ICSAB01	Iron	100000	99300	101	84400	114500	07/15/2025	13:43	LB136487
ICSA	Iron	101000	101000	100	85600	116500	07/15/2025	13:47	LB136487
ICSAB	Iron	96600	99300	97	84400	114500	07/15/2025	13:51	LB136487
ICSA01	Iron	100000	101000	99	85600	116500	07/16/2025	14:26	LB136511
ICSAB01	Iron	97800	99300	98	84400	114500	07/16/2025	14:38	LB136511
ICSA	Iron	100000	101000	99	85600	116500	07/16/2025	14:43	LB136511
ICSAB	Iron	97400	99300	98	84400	114500	07/16/2025	14:47	LB136511