

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2594
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073900.D **Time Analyzed:** 15:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.638	5.539	5.739	435.120	500.000	-13.0
Aroclor-1016-2	5.659	5.560	5.760	470.800	500.000	-5.8
Aroclor-1016-3	5.721	5.622	5.822	462.990	500.000	-7.4
Aroclor-1016-4	5.819	5.720	5.920	472.210	500.000	-5.6
Aroclor-1016-5	6.110	6.012	6.212	486.830	500.000	-2.6
Aroclor-1260-1	7.227	7.129	7.329	455.220	500.000	-9.0
Aroclor-1260-2	7.480	7.383	7.583	394.090	500.000	-21.2
Aroclor-1260-3	7.838	7.740	7.940	445.660	500.000	-10.9
Aroclor-1260-4	8.062	7.964	8.164	460.080	500.000	-8.0
Aroclor-1260-5	8.379	8.283	8.483	463.780	500.000	-7.2
Decachlorobiphenyl	10.172	10.076	10.276	48.620	50.000	-2.8
Tetrachloro-m-xylene	4.487	4.388	4.588	49.820	50.000	-0.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2594
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073900.D **Time Analyzed:** 15:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.854	4.758	4.958	510.520	500.000	2.1
Aroclor-1016-2	4.872	4.775	4.975	518.050	500.000	3.6
Aroclor-1016-3	5.048	4.951	5.151	527.740	500.000	5.5
Aroclor-1016-4	5.090	4.993	5.193	522.440	500.000	4.5
Aroclor-1016-5	5.303	5.207	5.407	560.430	500.000	12.1
Aroclor-1260-1	6.333	6.237	6.437	495.700	500.000	-0.9
Aroclor-1260-2	6.521	6.425	6.625	504.980	500.000	1.0
Aroclor-1260-3	6.673	6.577	6.777	488.090	500.000	-2.4
Aroclor-1260-4	7.142	7.046	7.246	509.160	500.000	1.8
Aroclor-1260-5	7.385	7.289	7.489	510.270	500.000	2.1
Decachlorobiphenyl	8.774	8.680	8.880	55.180	50.000	10.4
Tetrachloro-m-xylene	3.777	3.678	3.878	51.470	50.000	2.9