

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2594
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073915.D **Time Analyzed:** 20:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.635	5.539	5.739	437.240	500.000	-12.6
Aroclor-1016-2	5.656	5.560	5.760	466.120	500.000	-6.8
Aroclor-1016-3	5.719	5.622	5.822	461.790	500.000	-7.6
Aroclor-1016-4	5.816	5.720	5.920	470.970	500.000	-5.8
Aroclor-1016-5	6.108	6.012	6.212	489.050	500.000	-2.2
Aroclor-1260-1	7.225	7.129	7.329	468.890	500.000	-6.2
Aroclor-1260-2	7.478	7.383	7.583	403.420	500.000	-19.3
Aroclor-1260-3	7.836	7.740	7.940	450.340	500.000	-9.9
Aroclor-1260-4	8.060	7.964	8.164	459.360	500.000	-8.1
Aroclor-1260-5	8.377	8.283	8.483	466.750	500.000	-6.7
Decachlorobiphenyl	10.171	10.076	10.276	49.470	50.000	-1.1
Tetrachloro-m-xylene	4.484	4.388	4.588	49.580	50.000	-0.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENVI60
Lab Code: ACE **SDG NO.:** Q2594
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/17/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073915.D **Time Analyzed:** 20:08

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.854	4.758	4.958	505.390	500.000	1.1
Aroclor-1016-2	4.872	4.775	4.975	518.060	500.000	3.6
Aroclor-1016-3	5.048	4.951	5.151	518.280	500.000	3.7
Aroclor-1016-4	5.090	4.993	5.193	512.730	500.000	2.5
Aroclor-1016-5	5.303	5.207	5.407	567.650	500.000	13.5
Aroclor-1260-1	6.333	6.237	6.437	511.880	500.000	2.4
Aroclor-1260-2	6.521	6.425	6.625	526.500	500.000	5.3
Aroclor-1260-3	6.673	6.577	6.777	507.420	500.000	1.5
Aroclor-1260-4	7.142	7.046	7.246	520.360	500.000	4.1
Aroclor-1260-5	7.385	7.289	7.489	525.190	500.000	5.0
Decachlorobiphenyl	8.775	8.680	8.880	57.340	50.000	14.7
Tetrachloro-m-xylene	3.777	3.678	3.878	51.770	50.000	3.5