

Surrogate Summary

SDG No.: Q2594

Client: Environmental Restoration, LLC

Analytical Method: 608.3 PCB

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP073553.D	PIBLK-PP073553.D	Tetrachloro-m-xyl	1	20	16.2	81		70 (60)	130 (140)
		Decachlorobiphen	1	20	17.3	87		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	16.2	81		70 (60)	130 (140)
		Decachlorobiphen	2	20	17.1	85		70 (60)	130 (140)
I.BLK-PP073904.D	PIBLK-PP073904.D	Tetrachloro-m-xyl	1	20	18.6	93		70 (60)	130 (140)
		Decachlorobiphen	1	20	18.1	91		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	19.5	98		70 (60)	130 (140)
		Decachlorobiphen	2	20	20.9	104		70 (60)	130 (140)
PB168905BL	PB168905BL	Tetrachloro-m-xyl	1	20	21.4	107		70 (60)	130 (140)
		Decachlorobiphen	1	20	20.8	104		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	23.1	115		70 (60)	130 (140)
		Decachlorobiphen	2	20	24.1	121		70 (60)	130 (140)
PB168905BS	PB168905BS	Tetrachloro-m-xyl	1	20	21.4	107		70 (60)	130 (140)
		Decachlorobiphen	1	20	21.1	105		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	23.2	116		70 (60)	130 (140)
		Decachlorobiphen	2	20	24.0	120		70 (60)	130 (140)
PB168905BSD	PB168905BSD	Tetrachloro-m-xyl	1	20	21.2	106		70 (60)	130 (140)
		Decachlorobiphen	1	20	20.5	102		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	22.3	111		70 (60)	130 (140)
		Decachlorobiphen	2	20	24.1	121		70 (60)	130 (140)
Q2594-01	CC-071325-RW	Tetrachloro-m-xyl	1	20	19.8	99		70 (60)	130 (140)
		Decachlorobiphen	1	20	14.2	71		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	39.3	196	*	70 (60)	130 (140)
		Decachlorobiphen	2	20	15.2	76		70 (60)	130 (140)
I.BLK-PP073919.D	PIBLK-PP073919.D	Tetrachloro-m-xyl	1	20	18.1	91		70 (60)	130 (140)
		Decachlorobiphen	1	20	18.2	91		70 (60)	130 (140)
		Tetrachloro-m-xyl	2	20	18.8	94		70 (60)	130 (140)
		Decachlorobiphen	2	20	21.2	106		70 (60)	130 (140)