

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072125\
 Data File : PD089542.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jul 2025 13:16
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_D
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 22 03:52:21 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072125.M
 Quant Title : GC Extractables
 QLast Update : Tue Jul 22 03:49:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.879	141.9E6	949.4E6	50.000	50.000
28)	SA Decachlor...	9.071	8.070	196.0E6	1166.9E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.998	3.391	289.7E6	1466.5E6	50.000	50.000
3)	MA gamma-BHC...	4.328	3.727	275.8E6	1354.5E6	50.000	50.000
4)	MA Heptachlor	4.927	4.080	270.4E6	1359.6E6	50.000	50.000
5)	MB Aldrin	5.269	4.366	260.1E6	1333.0E6	50.000	50.000
6)	B beta-BHC	4.514	4.023	105.2E6	576.7E6	50.000	50.000
7)	B delta-BHC	4.763	4.260	266.7E6	1362.7E6	50.000	50.000
8)	B Heptachlo...	5.688	4.870	231.2E6	1200.8E6	50.000	50.000
9)	A Endosulfan I	6.072	5.244	217.6E6	1144.6E6	50.000	50.000
10)	B gamma-Chl...	5.944	5.123	233.0E6	1290.5E6	50.000	50.000
11)	B alpha-Chl...	6.024	5.187	232.9E6	1242.8E6	50.000	50.000
12)	B 4,4'-DDE	6.194	5.372	212.1E6	1266.0E6	50.000	50.000
13)	MA Dieldrin	6.345	5.510	233.6E6	1277.4E6	50.000	50.000
14)	MA Endrin	6.572	5.787	200.5E6	1163.4E6	50.000	50.000
15)	B Endosulfa...	6.784	6.078	197.0E6	1103.5E6	50.000	50.000
16)	A 4,4'-DDD	6.703	5.927	166.9E6	1054.6E6	50.000	50.000
17)	MA 4,4'-DDT	7.019	6.181	186.6E6	1143.3E6	50.000	50.000
18)	B Endrin al...	6.913	6.256	140.5E6	789.0E6	50.000	50.000
19)	B Endosulfa...	7.147	6.479	183.4E6	1062.3E6	50.000	50.000
20)	A Methoxychlor	7.492	6.751	97441869	586.7E6	50.000	50.000
21)	B Endrin ke...	7.628	6.989	196.9E6	1171.2E6	50.000	50.000
22)	Mirex	8.112	7.183	146.3E6	917.3E6	50.000	50.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD072125\
Data File : PD089542.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jul 2025 13:16
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_D
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 22 03:52:21 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD072125.M
Quant Title : GC Extractables
QLast Update : Tue Jul 22 03:49:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

