

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072125

Review By	Abdul	Review On	7/22/2025 8:03:16 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:33:20 AM
SubDirectory	PD072125	HP Acquire Method	HP Processing Method pd072125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24764
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089536.D	21 Jul 2025 11:27		AR\AJ	Ok
2	I.BLK	I.BLK	PD089537.D	21 Jul 2025 12:08		AR\AJ	Ok
3	PEM	PEM	PD089538.D	21 Jul 2025 12:22		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD089539.D	21 Jul 2025 12:35		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PD089540.D	21 Jul 2025 12:49		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD089541.D	21 Jul 2025 13:03		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD089542.D	21 Jul 2025 13:16		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD089543.D	21 Jul 2025 13:30		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD089544.D	21 Jul 2025 13:44		AR\AJ	Ok
10	PCHLORICC1000	PCHLORICC1000	PD089545.D	21 Jul 2025 13:57		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD089546.D	21 Jul 2025 14:11		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD089547.D	21 Jul 2025 14:25		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD089548.D	21 Jul 2025 14:38		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD089549.D	21 Jul 2025 14:52		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PD089550.D	21 Jul 2025 15:05		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD089551.D	21 Jul 2025 15:19		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD089552.D	21 Jul 2025 15:32		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD089553.D	21 Jul 2025 15:46		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072125

Review By	Abdul	Review On	7/22/2025 8:03:16 AM
Supervise By	mohammad	Supervise On	7/23/2025 1:33:20 AM
SubDirectory	PD072125	HP Acquire Method	HP Processing Method pd072125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24764		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD089554.D	21 Jul 2025 15:59		AR\AJ	Ok
20	PSTDICV050	ICVPD072125	PD089555.D	21 Jul 2025 16:25		AR\AJ	Ok
21	PCHLORICV500	ICVPD072125CHLOR	PD089556.D	21 Jul 2025 16:39		AR\AJ	Ok
22	PTOXICV500	ICVPD072125TOX	PD089557.D	21 Jul 2025 16:52		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072225

Review By	Abdul	Review On	7/23/2025 3:44:30 PM
Supervise By	mohammad	Supervise On	7/24/2025 1:26:10 AM
SubDirectory	PD072225	HP Acquire Method	HP Processing Method pd072125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24764
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089558.D	22 Jul 2025 08:19		AR\AJ	Ok
2	I.BLK	I.BLK	PD089559.D	22 Jul 2025 08:33		AR\AJ	Ok
3	PEM	PEM	PD089560.D	22 Jul 2025 08:46		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD089561.D	22 Jul 2025 09:31		AR\AJ	Ok
5	PB168867BL	PB168867BL	PD089562.D	22 Jul 2025 10:07		AR\AJ	Ok
6	PB168867BS	PB168867BS	PD089563.D	22 Jul 2025 10:20		AR\AJ	Ok
7	PB168803TB	PB168803TB	PD089564.D	22 Jul 2025 10:40		AR\AJ	Ok
8	Q2578-03	WC-A5-01-C	PD089565.D	22 Jul 2025 10:53		AR\AJ	Ok,M
9	Q2578-03MS	WC-A5-01-CMS	PD089566.D	22 Jul 2025 11:07		AR\AJ	Ok,M
10	Q2578-03MSD	WC-A5-01-CMSD	PD089567.D	22 Jul 2025 11:20		AR\AJ	Ok,M
11	Q2578-07	WC-A2-09-C	PD089568.D	22 Jul 2025 11:34		AR\AJ	Ok,M
12	Q2578-11	WC-A2-10-C	PD089569.D	22 Jul 2025 11:47		AR\AJ	Ok,M
13	Q2578-15	WC-A2-11-C	PD089570.D	22 Jul 2025 12:01		AR\AJ	Ok,M
14	Q2578-19	WC-A2-12-C	PD089571.D	22 Jul 2025 12:15		AR\AJ	Ok,M
15	Q2579-03	WC-A2-13-C	PD089572.D	22 Jul 2025 12:29		AR\AJ	Ok,M
16	Q2579-07	WC-A2-14-C	PD089573.D	22 Jul 2025 12:42		AR\AJ	Ok,M
17	I.BLK	I.BLK	PD089574.D	22 Jul 2025 12:56		AR\AJ	Ok
18	PSTDCCC050	PSTDCCC050	PD089575.D	22 Jul 2025 13:31		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072225

Review By	Abdul	Review On	7/23/2025 3:44:30 PM
Supervise By	mohammad	Supervise On	7/24/2025 1:26:10 AM
SubDirectory	PD072225	HP Acquire Method	HP Processing Method pd072125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24764		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB168906BL	PB168906BL	PD089576.D	22 Jul 2025 14:02		AR\AJ	Ok
20	PB168906BS	PB168906BS	PD089577.D	22 Jul 2025 14:16		AR\AJ	Ok
21	PB168906BSD	PB168906BSD	PD089578.D	22 Jul 2025 14:42		AR\AJ	Ok,M
22	Q2594-01	CC-071325-RW	PD089579.D	22 Jul 2025 14:56	DCB Low in 2nd column , TCMX low in 1st column	AR\AJ	ReRun
23	I.BLK	I.BLK	PD089580.D	22 Jul 2025 15:36		AR\AJ	Ok
24	PSTDCCC050	PSTDCCC050	PD089581.D	22 Jul 2025 15:49		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072825

Review By	yogesh	Review On	7/29/2025 7:28:43 AM
Supervise By	mohammad	Supervise On	7/29/2025 7:50:00 AM
SubDirectory	PD072825	HP Acquire Method	HP Processing Method pd072125 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24433,PP24651
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24764
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD089636.D	28 Jul 2025 09:40		AR\AJ	Ok
2	I.BLK	I.BLK	PD089637.D	28 Jul 2025 09:53		AR\AJ	Ok
3	PEM	PEM	PD089638.D	28 Jul 2025 10:07		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD089639.D	28 Jul 2025 11:28		AR\AJ	Ok
5	Q2594-01	CC-071325-RW	PD089640.D	28 Jul 2025 11:48		AR\AJ	Ok,M
6	PB168853BL	PB168853BL	PD089641.D	28 Jul 2025 12:07		AR\AJ	Ok
7	PB168853BS	PB168853BS	PD089642.D	28 Jul 2025 13:01		AR\AJ	Ok
8	Q2600-01	TRENCH	PD089643.D	28 Jul 2025 13:33		AR\AJ	Ok,M
9	Q2600-05	STOCK-PILE	PD089644.D	28 Jul 2025 13:46		AR\AJ	Ok,M
10	Q2600-05MS	STOCK-PILEMS	PD089645.D	28 Jul 2025 14:00	Comp#18 recovery fail	AR\AJ	Ok,M
11	Q2600-05MSD	STOCK-PILEMSD	PD089646.D	28 Jul 2025 14:13	Comp#18 recovery fail	AR\AJ	Ok,M
12	Q2600-09	END-OF-TRENCH	PD089647.D	28 Jul 2025 14:27		AR\AJ	Ok,M
13	PB169019BL	PB169019BL	PD089648.D	28 Jul 2025 14:40		AR\AJ	Ok
14	PB169019BS	PB169019BS	PD089649.D	28 Jul 2025 14:54		AR\AJ	Ok
15	I.BLK	I.BLK	PD089650.D	28 Jul 2025 15:07		AR\AJ	Ok
16	PSTDCCC050	PSTDCCC050	PD089651.D	28 Jul 2025 15:21		AR\AJ	Ok
17	Q2700-01	EO-03-072525	PD089652.D	28 Jul 2025 15:51		AR\AJ	Ok,M
18	Q2703-01	TP-4	PD089653.D	28 Jul 2025 16:05	DCB Low in 2nd column	AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD072825

Review By	yogesh	Review On	7/29/2025 7:28:43 AM
Supervise By	mohammad	Supervise On	7/29/2025 7:50:00 AM
SubDirectory	PD072825	HP Acquire Method	HP Processing Method pd072125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24433,PP24651		
Initial Calibration Stds	PP24744,PP24750,PP24751,PP24752,PP24753,PP24746,PP24755,PP24756,PP24757,PP24758,PP24748,PP24760,PP24761,PP24762,PP24763		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24764		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2706-01	RT-5417	PD089654.D	28 Jul 2025 16:18		AR\AJ	Ok,M
20	Q2706-03	ETGI-361	PD089655.D	28 Jul 2025 16:32		AR\AJ	Ok,M
21	Q2706-03MS	ETGI-361MS	PD089656.D	28 Jul 2025 16:46		AR\AJ	Ok,M
22	Q2706-03MSD	ETGI-361MSD	PD089657.D	28 Jul 2025 16:59		AR\AJ	Ok,M
23	I.BLK	I.BLK	PD089658.D	28 Jul 2025 18:36		AR\AJ	Ok
24	PSTDCCC050	PSTDCCC050	PD089659.D	28 Jul 2025 19:03	DCB high in 2nd column	AR\AJ	Ok
25	Q2481-14	CC0625-OXBL	PD089660.D	28 Jul 2025 20:25	END ccc out of tune , DCB low in both column , TCMX low in 2nd column	AR\AJ	Not Ok
26	Q2481-19	CC0627-CLOXAL	PD089661.D	28 Jul 2025 20:52	TCMX high in both column , DCB low in 2nd column , END ccc out of tune , F flag in TCMX	AR\AJ	Not Ok
27	Q248121	Q248121	PD089662.D	28 Jul 2025 21:20	END ccc out of tune , Typo , Need cleanup	AR\AJ	Not Ok
28	I.BLK	I.BLK	PD089663.D	28 Jul 2025 22:01		AR\AJ	Not Ok
29	PSTDCCC050	PSTDCCC050	PD089664.D	28 Jul 2025 22:14	Out of tune PEM tune time	AR\AJ	Not Ok

M : Manual Integration