

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2594

Analytical Method: 608.3 Pest

Client: Environmental Restoration, LLC

Datafile : PD089578.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168906BSD (Column 1)	alpha-BHC	0.0125	0.010	ug/L	80	3			37 (37)	140 (140)	20 (20)
	gamma-BHC (Lindane)	0.0125	0.011	ug/L	84	2			32 (32)	140 (140)	20 (20)
	Heptachlor	0.0125	0.011	ug/L	86	2			34 (34)	140 (140)	20 (20)
	Aldrin	0.0125	0.011	ug/L	87	1			42 (42)	140 (140)	20 (20)
	beta-BHC	0.0125	0.012	ug/L	93	1			17 (17)	147 (147)	20 (20)
	delta-BHC	0.0125	0.010	ug/L	83	4			19 (19)	140 (140)	20 (20)
	Heptachlor epoxide	0.0125	0.011	ug/L	90	2			37 (37)	142 (142)	20 (20)
	Endosulfan I	0.0125	0.011	ug/L	89	2			45 (45)	153 (153)	20 (20)
	gamma-Chlordane	0.0125	0.011	ug/L	88	2			45 (45)	140 (140)	20 (20)
	alpha-Chlordane	0.0125	0.011	ug/L	89	3			45 (45)	140 (140)	20 (20)
	4,4'-DDE	0.0125	0.011	ug/L	86	4			30 (30)	145 (145)	20 (20)
	Dieldrin	0.0125	0.011	ug/L	87	2			36 (36)	146 (146)	20 (20)
	Endrin	0.0125	0.011	ug/L	87	2			30 (30)	147 (147)	20 (20)
	Endosulfan II	0.0125	0.011	ug/L	91	0			1 (1)	202 (202)	20 (20)
	4,4'-DDD	0.0125	0.011	ug/L	87	2			31 (31)	141 (141)	20 (20)
	4,4'-DDT	0.0125	0.011	ug/L	88	2			25 (25)	160 (160)	20 (20)
	Endrin aldehyde	0.0125	0.012	ug/L	96	2			38 (38)	141 (141)	20 (20)
	Endosulfan sulfate	0.0125	0.011	ug/L	91	2			26 (26)	144 (144)	20 (20)
	Methoxychlor	0.0125	0.012	ug/L	96	3			30 (30)	151 (151)	20 (20)
	Endrin ketone	0.0125	0.011	ug/L	91	2			44 (44)	149 (149)	20 (20)
PB168906BSD (Column 2)	alpha-BHC	0.0125	0.012	ug/L	94	2			37 (37)	140 (140)	20 (20)
	gamma-BHC (Lindane)	0.0125	0.012	ug/L	94	2			32 (32)	140 (140)	20 (20)
	Heptachlor	0.0125	0.012	ug/L	96	1			34 (34)	140 (140)	20 (20)
	Aldrin	0.0125	0.012	ug/L	95	1			42 (42)	140 (140)	20 (20)
	beta-BHC	0.0125	0.012	ug/L	98	0			17 (17)	147 (147)	20 (20)
	delta-BHC	0.0125	0.012	ug/L	96	2			19 (19)	140 (140)	20 (20)
	Heptachlor epoxide	0.0125	0.012	ug/L	98	0			37 (37)	142 (142)	20 (20)
	Endosulfan I	0.0125	0.012	ug/L	98	3			45 (45)	153 (153)	20 (20)
	gamma-Chlordane	0.0125	0.012	ug/L	98	3			45 (45)	140 (140)	20 (20)
	alpha-Chlordane	0.0125	0.012	ug/L	98	2			45 (45)	140 (140)	20 (20)
	4,4'-DDE	0.0125	0.012	ug/L	98	3			30 (30)	145 (145)	20 (20)
	Dieldrin	0.0125	0.012	ug/L	97	2			36 (36)	146 (146)	20 (20)
	Endrin	0.0125	0.012	ug/L	98	2			30 (30)	147 (147)	20 (20)
	Endosulfan II	0.0125	0.012	ug/L	99	1			1 (1)	202 (202)	20 (20)
	4,4'-DDD	0.0125	0.012	ug/L	98	1			31 (31)	141 (141)	20 (20)
	4,4'-DDT	0.0125	0.012	ug/L	97	3			25 (25)	160 (160)	20 (20)
	Endrin aldehyde	0.0125	0.013	ug/L	103	1			38 (38)	141 (141)	20 (20)
	Endosulfan sulfate	0.0125	0.012	ug/L	99	1			26 (26)	144 (144)	20 (20)
	Methoxychlor	0.0125	0.012	ug/L	99	1			30 (30)	151 (151)	20 (20)
	Endrin ketone	0.0125	0.013	ug/L	100	2			44 (44)	149 (149)	20 (20)