

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: ENVI60

Lab Code: ACE

SDG No.: Q2594

Instrument ID: BNA_F

Calibration Date(s): 07/17/2025 07/17/2025

Calibration Time(s): 11:04 14:34

LAB FILE ID:		RRF2.5 = BF143140.D			RRF005 = BF143141.D		RRF010 = BF143142.D	
		RRF020 = BF143143.D			RRF040 = BF143144.D		RRF050 = BF143145.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.385	1.313	1.319	1.231	1.284	1.264	7.0
Benzaldehyde			1.167	1.163	1.443	1.411	1.193	18.1
Phenol-d6		1.733	1.640	1.657	1.554	1.632	1.591	6.7
Phenol		1.882	1.767	1.776	1.711	1.788	1.733	6.4
bis(2-Chloroethyl)ether		1.419	1.343	1.379	1.303	1.361	1.327	5.4
2-Chlorophenol		1.391	1.343	1.371	1.310	1.369	1.325	4.9
2-Methylphenol		1.178	1.111	1.141	1.097	1.160	1.114	4.6
2,2-oxybis(1-Chloropropane)		2.670	2.553	2.545	2.421	2.517	2.461	6.5
Acetophenone		0.527	0.504	0.501	0.455	0.476	0.474	8.6
3+4-Methylphenols			1.467	1.469	1.341	1.402	1.351	8.9
n-Nitroso-di-n-propylamine	1.101	1.111	1.051	1.028	0.986	1.023	1.021	6.8
Nitrobenzene-d5		0.407	0.394	0.417	0.399	0.417	0.401	3.5
Hexachloroethane		0.522	0.498	0.522	0.495	0.525	0.502	4.7
Nitrobenzene		0.380	0.370	0.388	0.375	0.388	0.374	3.9
Isophorone		0.748	0.707	0.711	0.692	0.734	0.708	3.8
2-Nitrophenol		0.133	0.141	0.164	0.172	0.182	0.162	11.2
2,4-Dimethylphenol		0.356	0.340	0.341	0.324	0.338	0.331	5.7
bis(2-Chloroethoxy)methane		0.463	0.437	0.442	0.413	0.430	0.425	6.3
2,4-Dichlorophenol		0.292	0.282	0.287	0.276	0.289	0.279	4.4
Naphthalene		1.118	1.038	1.032	0.954	0.986	0.985	8.9
4-Chloroaniline		0.437	0.417	0.419	0.390	0.408	0.402	6.3
Hexachlorobutadiene		0.193	0.190	0.191	0.181	0.188	0.184	4.6
Caprolactam			0.078	0.084	0.087	0.091	0.084	5.3
4-Chloro-3-methylphenol		0.319	0.310	0.307	0.301	0.311	0.303	4.0
2-Methylnaphthalene		0.667	0.634	0.628	0.580	0.603	0.599	8.2
Hexachlorocyclopentadiene			0.337	0.373	0.373	0.404	0.376	6.1
2,4,6-Trichlorophenol		0.391	0.406	0.412	0.384	0.429	0.400	4.6
2-Fluorobiphenyl		1.761	1.644	1.547	1.358	1.414	1.462	13.6
2,4,5-Trichlorophenol		0.404	0.387	0.411	0.405	0.411	0.400	2.8
1,1-Biphenyl		1.748	1.667	1.640	1.489	1.553	1.560	8.6
2-Chloronaphthalene		1.268	1.223	1.192	1.103	1.166	1.155	7.1
2-Nitroaniline		0.305	0.333	0.372	0.377	0.396	0.362	8.9
Dimethylphthalate		1.422	1.340	1.340	1.279	1.325	1.304	6.2
Acenaphthylene		2.134	2.052	2.011	1.860	1.958	1.935	7.7

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: ENVI60

Lab Code: ACE

SDG No.: Q2594

Instrument ID: BNA_F

Calibration Date(s): 07/17/2025 07/17/2025

Calibration Time(s): 11:04 14:34

LAB FILE ID:		RRF2.5 = BF143140.D			RRF005 = BF143141.D		RRF010 = BF143142.D	
		RRF020 = BF143143.D			RRF040 = BF143144.D		RRF050 = BF143145.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.225	0.250	0.272	0.278	0.289	0.265	8.1
3-Nitroaniline		0.285	0.300	0.315	0.317	0.334	0.310	5.1
Acenaphthene		1.278	1.194	1.184	1.096	1.155	1.144	7.5
2,4-Dinitrophenol			0.051	0.081	0.104	0.116	0.098	28.0
4-Nitrophenol			0.200	0.228	0.246	0.250	0.232	7.7
Dibenzofuran		1.929	1.807	1.766	1.633	1.698	1.698	9.0
2,4-Dinitrotoluene		0.269	0.296	0.342	0.354	0.372	0.333	11.0
Diethylphthalate		1.377	1.322	1.335	1.269	1.329	1.289	5.8
4-Chlorophenyl-phenylether		0.710	0.670	0.657	0.612	0.633	0.635	7.9
Fluorene		1.480	1.402	1.322	1.212	1.257	1.274	10.8
4-Nitroaniline		0.236	0.255	0.266	0.280	0.289	0.266	6.4
4,6-Dinitro-2-methylphenol			0.054	0.087	0.097	0.110	0.094	22.9
n-Nitrosodiphenylamine		0.761	0.728	0.721	0.674	0.726	0.704	6.0
2,4,6-Tribromophenol		0.200	0.202	0.212	0.210	0.218	0.207	3.5
4-Bromophenyl-phenylether		0.252	0.238	0.241	0.228	0.247	0.238	4.3
Hexachlorobenzene		0.269	0.248	0.252	0.239	0.258	0.249	4.8
Atrazine		0.184	0.187	0.198	0.196	0.208	0.194	4.3
Pentachlorophenol			0.119	0.146	0.148	0.160	0.147	9.7
Phenanthrene		1.191	1.100	1.117	1.020	1.069	1.062	8.0
Anthracene		1.184	1.136	1.134	1.038	1.095	1.083	7.1
Carbazole		1.042	0.969	0.982	0.929	0.963	0.946	6.9
Di-n-butylphthalate		1.076	1.054	1.129	1.074	1.119	1.070	4.4
Fluoranthene		1.085	0.987	1.018	0.953	0.979	0.975	6.9
Pyrene		1.909	1.783	1.783	1.715	1.740	1.707	8.7
Terphenyl-d14		1.567	1.456	1.419	1.311	1.342	1.344	11.4
Butylbenzylphthalate		0.438	0.472	0.536	0.544	0.582	0.523	9.6
3,3-Dichlorobenzidine			0.437	0.495	0.429	0.476	0.446	7.6
Benzo (a) anthracene		1.374	1.367	1.381	1.282	1.422	1.339	4.9
Chrysene		1.282	1.174	1.247	1.206	1.218	1.204	4.3
Bis (2-ethylhexyl) phthalate		0.691	0.755	0.824	0.789	0.862	0.791	7.0
Di-n-octyl phthalate			1.274	1.428	1.384	1.550	1.434	6.8
Benzo (b) fluoranthene		1.231	1.136	1.287	1.147	1.352	1.222	6.3
Benzo (k) fluoranthene		1.168	1.152	1.093	1.134	1.069	1.090	7.1
Benzo (a) pyrene		1.129	1.102	1.147	1.121	1.192	1.126	3.4
Indeno (1,2,3-cd) pyrene		1.424	1.336	1.430	1.413	1.505	1.410	3.9

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: ENVI60

Lab Code: ACE

SDG No.: Q2594

Instrument ID: BNA_F

Calibration Date(s): 07/17/2025 07/17/2025

Calibration Time(s): 11:04 14:34

LAB FILE ID:		RRF2.5 = BF143140.D		RRF005 = BF143141.D		RRF010 = BF143142.D		
		RRF020 = BF143143.D		RRF040 = BF143144.D		RRF050 = BF143145.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		1.159	1.103	1.179	1.147	1.221	1.148	3.9
Benzo (g,h,i) perylene		1.100	1.047	1.140	1.109	1.196	1.110	4.4
1,2,4,5-Tetrachlorobenzene		0.638	0.598	0.601	0.550	0.582	0.577	6.9
1,4-Dioxane		0.622	0.582	0.614	0.600	0.625	0.598	4.1
2,3,4,6-Tetrachlorophenol		0.321	0.331	0.345	0.333	0.351	0.331	4.3