

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF071825

Review By	Rahul	Review On	7/21/2025 8:15:21 AM
Supervise By	Jagrut	Supervise On	7/21/2025 9:31:31 AM
SubDirectory	BF071825	HP Acquire Method	BNA_F
		HP Processing Method	bf071725

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12674,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143158.D	18 Jul 2025 10:55		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143159.D	18 Jul 2025 11:24		RC/JU	Ok,M
3	PB168737BL	PB168737BL	BF143160.D	18 Jul 2025 11:54		RC/JU	Ok
4	PB168854BS	PB168854BS	BF143161.D	18 Jul 2025 12:23	Already analyzed	RC/JU	Not Ok
5	PB168904BS	PB168904BS	BF143162.D	18 Jul 2025 12:53		RC/JU	Ok,M
6	PB168904BSD	PB168904BSD	BF143163.D	18 Jul 2025 13:22		RC/JU	Ok,M
7	PB168904BL	PB168904BL	BF143164.D	18 Jul 2025 13:52		RC/JU	Ok
8	Q2585-04	EFF-WW	BF143165.D	18 Jul 2025 14:25	Surrogate and Internal standard fail	RC/JU	Ok
9	Q2594-01	CC-071325-RW	BF143166.D	18 Jul 2025 14:54	Surrogate fail.	RC/JU	Ok

M : Manual Integration