



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

T&A Construction Inc

Project Name: Kingsland Point Park Water Main

Project # N/A

Order ID # Q2600

Test Name: Cyanide,Hexavalent Chromium,Ignitability,pH

A. Number of Samples and Date of Receipt:

12 Solid samples were received on 07/14/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Cyanide,Hexavalent Chromium,Ignitability,pH. This data package contains results for Cyanide,Hexavalent Chromium,Ignitability,pH.

C. Analytical Techniques:

The analysis of Ignitability was based on method 1030, The analysis of Hexavalent Chromium was based on method 7196A, The analysis of Cyanide was based on method 9012B and The analysis of pH was based on method 9045D.

D. QA/ QC Samples:

The Holding Times were met for all samples except for END-OF-TRENCH of pH, for STOCK-PILE of pH and for TRENCH of pH as samples were receive out of holding time.

The Blank Spike met requirements for all compounds.

The Duplicate (END-OF-TRENCHDUP) analysis met criteria for all compounds except for Cyanide due to the results are below Reporting limit.

The Matrix Spike analysis met criteria for all compounds.

The Matrix Spike Duplicate analysis met criteria for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____