



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

T&A Construction Inc

Project Name: Kingsland Point Park Water Main

Project # N/A

Order ID # Q2600

Test Name: Mercury, Metals ICP-TAL

A. Number of Samples and Date of Receipt:

12 Solid samples were received on 07/14/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Cyanide, EPH, Gasoline Range Organics, Hexavalent Chromium, Ignitability, Mercury, Metals ICP-TAL, PCB, Pesticide-TCL, pH, SVOC-TCL BNA -20, TCL+30/TAL, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL and VOC-TCLVOA-10. This data package contains results for Mercury, Metals ICP-TAL.

C. Analytical Techniques:

The analysis of Metals ICP-TAL was based on method 6010D, digestion based on method 3050 (soils). The analysis and digestion of Mercury was based on method 7471B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all compounds.

The Duplicate analysis met criteria for all compounds.

The Matrix Spike (NB-07-071125MS) analysis met criteria for all compounds except for Antimony, Copper, Mercury, Silver and Zinc due to Chemical Interference during Digestion Process.

The Matrix Spike Duplicate (NB-07-071125MSD) analysis met criteria for all compounds except for Antimony, Copper, Mercury and Silver due to Chemical Interference during Digestion Process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution (NB-07-071125L) met criteria for all compounds except for Iron due to sample matrix interference.

E. Additional Comments:

The Post Digest Spike (NB-07-071125A) analysis met criteria for all compounds except for Zinc due to unknown chemical interference of matrix with the addition of spike amount after digestion and before analysis; matrix has suppression effect during addition of spike.



I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature _____