

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB136472

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/15/2025	08:40
2	CAL2	1	Water	NA	NA	20.2	7.00	07/15/2025	08:41
3	CAL3	1	Water	NA	NA	20.2	10.02	07/15/2025	08:42
4	ICV	1	Water	NA	NA	20.2	7.01	07/15/2025	08:45
5	CCV1	1	Water	NA	NA	20.2	2.01	07/15/2025	08:47
6	Q2592-01	1	Solid	20.02	20	22.8	6.94	07/15/2025	09:00
7	Q2592-01DUP	1	Solid	20.03	20	22.9	6.96	07/15/2025	09:02
8	Q2600-01	1	Solid	20.02	20	24.6	7.55	07/15/2025	09:15
9	Q2600-05	1	Solid	20.03	20	24.4	7.68	07/15/2025	09:25
10	Q2600-09	1	Solid	20.02	20	24.1	7.24	07/15/2025	09:30
11	CCV2	1	Water	NA	NA	20.2	12.02	07/15/2025	09:33

WORKLIST(Hardcopy Internal Chain)

136472

WorkList Name : pbs q2600 WorkList ID : 190717 Department : Wet-Chemistry Date : 07-15-2025 08:20:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2592-01	WC-SOIL-20250711	Solid	pH	Cool 4 deg C	PARS02	D51	07/11/2025	9045D
Q2600-01	TRENCH	Solid	pH	Cool 4 deg C	TACO01	D41	07/14/2025	9045D
Q2600-05	STOCK-PILE	Solid	pH	Cool 4 deg C	TACO01	D41	07/14/2025	9045D
Q2600-09	END-OF-TRENCH	Solid	pH	Cool 4 deg C	TACO01	D41	07/14/2025	9045D

Date/Time 07/15/25 08:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 07/15/25 13:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]