

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: T&A Construction Inc

SDG No.: Q2600

Contract: TACO01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168846BS							
Aluminum	mg/Kg	100	101		101	80 - 120	P
Antimony	mg/Kg	40.0	40.1		100	80 - 120	P
Arsenic	mg/Kg	40.0	35.5		89	80 - 120	P
Barium	mg/Kg	10.0	10.2		102	80 - 120	P
Beryllium	mg/Kg	10.0	10.3		103	80 - 120	P
Cadmium	mg/Kg	10.0	9.35		94	80 - 120	P
Calcium	mg/Kg	50.0	50.2	J	100	80 - 120	P
Chromium	mg/Kg	20.0	19.6		98	80 - 120	P
Cobalt	mg/Kg	10.0	9.66		97	80 - 120	P
Copper	mg/Kg	15.0	16.0		107	80 - 120	P
Iron	mg/Kg	150	162		108	80 - 120	P
Lead	mg/Kg	50.0	46.8		94	80 - 120	P
Magnesium	mg/Kg	100	101		101	80 - 120	P
Manganese	mg/Kg	10.0	9.98		100	80 - 120	P
Nickel	mg/Kg	25.0	24.3		97	80 - 120	P
Potassium	mg/Kg	500	507		101	80 - 120	P
Selenium	mg/Kg	100	97.2		97	80 - 120	P
Silver	mg/Kg	3.8	3.68		97	80 - 120	P
Sodium	mg/Kg	150	137		91	80 - 120	P
Thallium	mg/Kg	100	98.8		99	80 - 120	P
Vanadium	mg/Kg	15.0	14.6		97	80 - 120	P
Zinc	mg/Kg	10.0	10.4		104	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: T&A Construction Inc

SDG No.: Q2600

Contract: TACO01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168891BS Mercury	mg/Kg	0.24	0.24		100	80 - 120	CV