

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: T&A Construction Inc

SDG No.: Q2600

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
pb168847TB		WATER		Batch Number:	PB168893		Prep Date:	07/16/2025	
	Mercury	0.76	<2	U	2.00	CV	07/17/2025	11:38	LB136516
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168893BL		WATER		Batch Number:	PB168893		Prep Date:	07/16/2025	
	Mercury	0.094	<0.2	J	0.20	CV	07/17/2025	10:41	LB136516

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: T&A Construction Inc

SDG No.: Q2600

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168847TB		WATER		Batch Number:	PB168888		Prep Date:	07/16/2025	
	Arsenic	25.6	<50	U	100	P	07/17/2025	16:50	LB136533
	Barium	72.8	<250	U	500	P	07/17/2025	16:50	LB136533
	Cadmium	2.50	<15	U	30.0	P	07/17/2025	16:50	LB136533
	Chromium	10.6	<25	U	50.0	P	07/17/2025	16:50	LB136533
	Lead	11.5	<30	U	60.0	P	07/17/2025	16:50	LB136533
	Selenium	48.2	<50	U	100	P	07/17/2025	16:50	LB136533
	Silver	8.10	<25	U	50.0	P	07/17/2025	16:50	LB136533
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168888BL		WATER		Batch Number:	PB168888		Prep Date:	07/16/2025	
	Arsenic	25.6	<50	U	100	P	07/17/2025	16:33	LB136533
	Barium	72.8	<250	U	500	P	07/17/2025	16:33	LB136533
	Cadmium	2.50	<15	U	30.0	P	07/17/2025	16:33	LB136533
	Chromium	10.6	<25	U	50.0	P	07/17/2025	16:33	LB136533
	Lead	11.5	<30	U	60.0	P	07/17/2025	16:33	LB136533
	Selenium	48.2	<50	U	100	P	07/17/2025	16:33	LB136533
	Silver	8.10	<25	U	50.0	P	07/17/2025	16:33	LB136533