

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: T&A Construction Inc **SDG No.:** Q2600
Contract: TACO01 **Lab Code:** ACE
Matrix: Water **Level:** LOW **Client ID:** HR-MCN-COMP-01A
Sample ID: Q2614-06 **Spiked ID:** Q2614-06A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	42.6		2.00	U	40.0	106		CV