

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: T&A Construction Inc **Level:** LOW **SDG No.:** Q2600
Contract: TACO01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2614-06 **Client ID:** HR-MCN-COMP-01DUP
Percent Solids for Sample: NA **Duplicate ID** Q2614-06DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	777		795		2		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	60.0	U	11.8	J	200.0		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: T&A Construction Inc **Level:** LOW **SDG No.:** Q2600
Contract: TACO01 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q2614-06MS **Client ID:** HR-MCN-COMP-01MSD
Percent Solids for Sample: NA **Duplicate ID** Q2614-06MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4090		4040		1		P
Barium	ug/L	20	2000		1930		4		P
Cadmium	ug/L	20	964		945		2		P
Chromium	ug/L	20	1800		1780		1		P
Lead	ug/L	20	4620		4520		2		P
Mercury	ug/L	20	26.6		31.5		17		CV
Selenium	ug/L	20	10400		10200		2		P
Silver	ug/L	20	378		371		2		P