

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: T&A Construction Inc

SDG No.: Q2600

Contract: TACO01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168888BS							
Arsenic	ug/L	4000	4070		102	80 - 120	P
Barium	ug/L	1000	1050		105	80 - 120	P
Cadmium	ug/L	1000	968		97	80 - 120	P
Chromium	ug/L	2000	1850		92	80 - 120	P
Lead	ug/L	5000	4810		96	80 - 120	P
Selenium	ug/L	10000	10400		104	80 - 120	P
Silver	ug/L	380	384		101	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: T&A Construction Inc

SDG No.: Q2600

Contract: TACO01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168893BS Mercury	ug/L	4.0	4.48		112	80 - 120	CV