

Continuing Calibration Report for SequenceID : FC071625AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069458.D

Aliphatic C9-C12	8429055.000	60.000	3.297	6.597	140484.250	138999.501	-1.068
Aliphatic C12-C16	5860780.000	40.000	6.598	10.000	146519.500	144073.407	-1.698
Aliphatic C16-C21	8127643.000	60.000	10.001	13.371	135460.717	135996.946	0.394
Aliphatic C21-C28	9303508.000	80.000	13.372	17.037	116293.850	118989.055	2.265
Aliphatic C28-C40	10489112.000	120.000	17.038	22.012	87409.267	89883.162	2.752
Aliphatic EPH	42210098.000	360.000	3.297	22.012	117250.272	118243.964	0.840