

Continuing Calibration Report for SequenceID : FC071625AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069469.D

Aliphatic C9-C12	7802791.000	60.000	3.297	6.597	130046.517	138999.501	6.441
Aliphatic C12-C16	5268638.000	40.000	6.598	10.000	131715.950	144073.407	8.577
Aliphatic C16-C21	7395694.000	60.000	10.001	13.371	123261.567	135996.946	9.364
Aliphatic C21-C28	8615350.000	80.000	13.372	17.037	107691.875	118989.055	9.494
Aliphatic C28-C40	9784801.000	120.000	17.038	22.012	81540.008	89883.162	9.282
Aliphatic EPH	38867274.000	360.000	3.297	22.012	107964.650	118243.964	8.693