

Continuing Calibration Report for SequenceID : FC071625AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069476.D

Aliphatic C9-C12	7813680.000	60.000	3.297	6.597	130228.000	138999.501	6.310
Aliphatic C12-C16	5274676.000	40.000	6.598	10.000	131866.900	144073.407	8.472
Aliphatic C16-C21	7394008.000	60.000	10.001	13.371	123233.467	135996.946	9.385
Aliphatic C21-C28	8600619.000	80.000	13.372	17.037	107507.738	118989.055	9.649
Aliphatic C28-C40	9724515.000	120.000	17.038	22.012	81037.625	89883.162	9.841
Aliphatic EPH	38807498.000	360.000	3.297	22.012	107798.606	118243.964	8.834