

Continuing Calibration Report for SequenceID : FC071725AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC069485.D

|                   |              |         |        |        |            |            |       |
|-------------------|--------------|---------|--------|--------|------------|------------|-------|
| Aliphatic C9-C12  | 7788678.000  | 60.000  | 3.303  | 6.599  | 129811.300 | 138999.501 | 6.610 |
| Aliphatic C12-C16 | 5258194.000  | 40.000  | 6.600  | 10.002 | 131454.850 | 144073.407 | 8.758 |
| Aliphatic C16-C21 | 7374441.000  | 60.000  | 10.003 | 13.370 | 122907.350 | 135996.946 | 9.625 |
| Aliphatic C21-C28 | 8646336.000  | 80.000  | 13.371 | 17.036 | 108079.200 | 118989.055 | 9.169 |
| Aliphatic C28-C40 | 9879171.000  | 120.000 | 17.037 | 22.013 | 82326.425  | 89883.162  | 8.407 |
| Aliphatic EPH     | 38946820.000 | 360.000 | 3.303  | 22.013 | 108185.611 | 118243.964 | 8.506 |