

Continuing Calibration Report for SequenceID : FD071625AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049548.D

Aromatic C10-C12	5757629.000	40.000	4.425	6.169	143940.725	137773.911	-4.476
Aromatic C12-C16	9750119.000	60.000	6.170	8.797	162501.983	156307.806	-3.963
Aromatic C16-C21	12946992.000	80.000	8.798	13.086	161837.400	161283.467	-0.343
Aromatic C21-C36	26398938.000	180.000	13.087	18.506	146660.767	140022.851	-4.741
Aromatic EPH	54853678.000	360.000	4.425	18.506	152371.328	147211.709	-3.505