

Continuing Calibration Report for SequenceID : FD071625AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049559.D

Aromatic C10-C12	5613030.000	40.000	4.425	6.169	140325.750	137773.911	-1.852
Aromatic C12-C16	9286635.000	60.000	6.170	8.797	154777.250	156307.806	0.979
Aromatic C16-C21	12763503.000	80.000	8.798	13.086	159543.788	161283.467	1.079
Aromatic C21-C36	24714882.000	180.000	13.087	18.506	137304.900	140022.851	1.941
Aromatic EPH	52378050.000	360.000	4.425	18.506	145494.583	147211.709	1.166