

Continuing Calibration Report for SequenceID : FD071625AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049566.D

Aromatic C10-C12	5768561.000	40.000	4.425	6.169	144214.025	137773.911	-4.674
Aromatic C12-C16	9609520.000	60.000	6.170	8.797	160158.667	156307.806	-2.464
Aromatic C16-C21	13064098.000	80.000	8.798	13.086	163301.225	161283.467	-1.251
Aromatic C21-C36	25127041.000	180.000	13.087	18.506	139594.672	140022.851	0.306
Aromatic EPH	53569220.000	360.000	4.425	18.506	148803.389	147211.709	-1.081