

Continuing Calibration Report for SequenceID : FD071725AR

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FD049569.D

|                  |              |         |        |        |            |            |        |
|------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aromatic C10-C12 | 5740395.000  | 40.000  | 4.430  | 6.172  | 143509.875 | 137773.911 | -4.163 |
| Aromatic C12-C16 | 9483258.000  | 60.000  | 6.173  | 8.799  | 158054.300 | 156307.806 | -1.117 |
| Aromatic C16-C21 | 12984701.000 | 80.000  | 8.800  | 13.090 | 162308.763 | 161283.467 | -0.636 |
| Aromatic C21-C36 | 25300563.000 | 180.000 | 13.091 | 18.508 | 140558.683 | 140022.851 | -0.383 |
| Aromatic EPH     | 53508917.000 | 360.000 | 4.430  | 18.508 | 148635.881 | 147211.709 | -0.967 |