

Continuing Calibration Report for SequenceID : FD071725AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049576.D

Aromatic C10-C12	5912578.000	40.000	4.430	6.172	147814.450	137773.911	-7.288
Aromatic C12-C16	9954029.000	60.000	6.173	8.799	165900.483	156307.806	-6.137
Aromatic C16-C21	13574205.000	80.000	8.800	13.090	169677.563	161283.467	-5.205
Aromatic C21-C36	26138197.000	180.000	13.091	18.508	145212.206	140022.851	-3.706
Aromatic EPH	55579009.000	360.000	4.430	18.508	154386.136	147211.709	-4.874