

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB071425

Review By	yogesh	Review On	7/14/2025 1:38:55 PM
Supervise By	mohammad	Supervise On	7/16/2025 1:42:50 AM
SubDirectory	FB071425	HP Acquire Method	HP Processing Method
			FB071425
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24717,PP24720,PP24721,PP24722,PP24723,PP24724		
CCC	PP24726,PP24727		
Internal Standard/PEM			
ICV/I.BLK	PP24718,PP24725		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB032098.D	14 Jul 2025 9:57		YP/AJ	Ok,M
2	10 GRO STD		FB032099.D	14 Jul 2025 10:50		YP/AJ	Ok
3	20 GRO STD		FB032100.D	14 Jul 2025 11:18		YP/AJ	Ok
4	50 GRO STD		FB032101.D	14 Jul 2025 11:46		YP/AJ	Ok
5	100 GRO STD		FB032102.D	14 Jul 2025 13:19		YP/AJ	Ok
6	FB071425GROICV		FB032103.D	14 Jul 2025 14:14		YP/AJ	Ok
7	20 PPB GRO STD		FB032104.D	14 Jul 2025 14:55		YP/AJ	Ok
8	VBF0714S1		FB032105.D	14 Jul 2025 15:40		YP/AJ	Ok
9	VBF0714S2		FB032106.D	14 Jul 2025 16:08		YP/AJ	Ok,M
10	BSF0714S1		FB032107.D	14 Jul 2025 16:35		YP/AJ	Ok
11	Q2589-01		FB032108.D	14 Jul 2025 17:03	Vial-A, Surr FAl	YP/AJ	Not Ok
12	Q2589-01		FB032109.D	14 Jul 2025 17:31	Vial-B, Not purged	YP/AJ	Not Ok
13	Q2589-01		FB032110.D	14 Jul 2025 17:59	Vial-C	YP/AJ	Ok
14	BSF0714S2		FB032111.D	14 Jul 2025 18:26		YP/AJ	Ok
15	20 PPB GRO STD		FB032112.D	14 Jul 2025 18:54		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB071625

Review By	yogesh	Review On	7/16/2025 1:39:40 PM
Supervise By	mohammad	Supervise On	7/18/2025 3:11:08 AM
SubDirectory	FB071625	HP Acquire Method	HP Processing Method
			FB071425
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24717,PP24720,PP24721,PP24722,PP24723,PP24724		
CCC Internal Standard/PEM	PP24732,PP24733,PP24734		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24718,PP24725		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB032131.D	16 Jul 2025 9:31		YP/AJ	Ok
2	VBF0716S1		FB032132.D	16 Jul 2025 10:13		YP/AJ	Ok
3	VBF0716S2		FB032133.D	16 Jul 2025 10:41		YP/AJ	Ok
4	BSF0716S1		FB032134.D	16 Jul 2025 11:08		YP/AJ	Ok
5	Q2600-09		FB032135.D	16 Jul 2025 12:02	Vial-C, Need Dilution	YP/AJ	Dilution
6	Q2600-05		FB032136.D	16 Jul 2025 12:29	Vial-C, Need Dilution	YP/AJ	Dilution
7	Q2600-01		FB032137.D	16 Jul 2025 12:57	Vial-C	YP/AJ	Ok
8	BSF0716S2		FB032138.D	16 Jul 2025 13:25		YP/AJ	Ok
9	20 PPB GRO STD		FB032139.D	16 Jul 2025 13:52		YP/AJ	Ok
10	Q2600-09		FB032140.D	16 Jul 2025 14:38	Vial-C	YP/AJ	Ok
11	Q2600-05		FB032141.D	16 Jul 2025 15:34	Vial-C	YP/AJ	Ok,M
12	I.BLK		FB032142.D	16 Jul 2025 16:01		YP/AJ	Ok
13	Q2600-05		FB032143.D	16 Jul 2025 16:42	Vial-C ,no required	YP/AJ	Not Ok
14	BSF0716S3		FB032144.D	16 Jul 2025 17:10		YP/AJ	Ok
15	20 PPB GRO STD		FB032145.D	16 Jul 2025 17:37		YP/AJ	Ok

M : Manual Integration