

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073834.D **Time Analyzed:** 04:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.637	5.539	5.739	406.110	500.000	-18.8
Aroclor-1016-2	5.658	5.560	5.760	444.870	500.000	-11.0
Aroclor-1016-3	5.721	5.622	5.822	434.180	500.000	-13.2
Aroclor-1016-4	5.818	5.720	5.920	441.020	500.000	-11.8
Aroclor-1016-5	6.110	6.012	6.212	459.000	500.000	-8.2
Aroclor-1260-1	7.227	7.129	7.329	442.950	500.000	-11.4
Aroclor-1260-2	7.478	7.383	7.583	388.470	500.000	-22.3
Aroclor-1260-3	7.838	7.740	7.940	428.260	500.000	-14.3
Aroclor-1260-4	8.062	7.964	8.164	439.070	500.000	-12.2
Aroclor-1260-5	8.379	8.283	8.483	434.140	500.000	-13.2
Decachlorobiphenyl	10.171	10.076	10.276	46.520	50.000	-7.0
Tetrachloro-m-xylene	4.486	4.388	4.588	45.710	50.000	-8.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/16/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073834.D **Time Analyzed:** 04:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.855	4.758	4.958	468.590	500.000	-6.3
Aroclor-1016-2	4.872	4.775	4.975	472.680	500.000	-5.5
Aroclor-1016-3	5.049	4.951	5.151	477.370	500.000	-4.5
Aroclor-1016-4	5.090	4.993	5.193	474.450	500.000	-5.1
Aroclor-1016-5	5.303	5.207	5.407	494.620	500.000	-1.1
Aroclor-1260-1	6.333	6.237	6.437	483.280	500.000	-3.3
Aroclor-1260-2	6.522	6.425	6.625	491.070	500.000	-1.8
Aroclor-1260-3	6.673	6.577	6.777	469.470	500.000	-6.1
Aroclor-1260-4	7.142	7.046	7.246	479.030	500.000	-4.2
Aroclor-1260-5	7.385	7.289	7.489	473.830	500.000	-5.2
Decachlorobiphenyl	8.774	8.680	8.880	53.240	50.000	6.5
Tetrachloro-m-xylene	3.777	3.678	3.878	46.020	50.000	-8.0