

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2600</u>	Analytical Method:	<u>8082A</u>
Client:	<u>T&A Construction Inc</u>	DataFile :	<u>PP073826.D</u>

		Sample				Rec		RPD		Limits	
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2598-01MSD	Client Sample ID:		OK-01-071425MSD							
(Column 1)											
AR1016	172.1	0	162	ug/kg	94		4		55	146	15
AR1260	172.1	0	150	ug/kg	87		5		54	119	15
Lab Sample ID:	Q2598-01MSD	Client Sample ID:		OK-01-071425MSD							
(Column 2)											
AR1016	172.1	0	159	ug/kg	92		7		55	146	15
AR1260	172.1	0	146	ug/kg	85		7		54	119	15