

Surrogate Summary

SDG No.: **Q2600**

Client: **T&A Construction Inc**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD089537.D	PIBLK-PD089537.D	Decachlorobiphen	1	20	21.2	106		57	171
		Tetrachloro-m-xyl	1	20	18.5	92		61	148
		Decachlorobiphen	2	20	20.5	103		57	171
		Tetrachloro-m-xyl	2	20	19.9	100		61	148
I.BLK-PD089637.D	PIBLK-PD089637.D	Decachlorobiphen	1	20	17.9	90		57	171
		Tetrachloro-m-xyl	1	20	15.1	76		61	148
		Decachlorobiphen	2	20	15.8	79		57	171
		Tetrachloro-m-xyl	2	20	14.6	73		61	148
PB168853BL	PB168853BL	Decachlorobiphen	1	20	15.5	77		20	144
		Tetrachloro-m-xyl	1	20	13.1	66		19	148
		Decachlorobiphen	2	20	12.9	64		20	144
		Tetrachloro-m-xyl	2	20	13.3	66		19	148
PB168853BS	PB168853BS	Decachlorobiphen	1	20	17.2	86		20	144
		Tetrachloro-m-xyl	1	20	14.3	72		19	148
		Decachlorobiphen	2	20	15.2	76		20	144
		Tetrachloro-m-xyl	2	20	14.2	71		19	148
Q2600-01	TRENCH	Decachlorobiphen	1	20	10.7	53		20	144
		Tetrachloro-m-xyl	1	20	8.39	42		19	148
		Decachlorobiphen	2	20	6.60	33		20	144
		Tetrachloro-m-xyl	2	20	7.54	38		19	148
Q2600-05	STOCK-PILE	Decachlorobiphen	1	20	11.6	58		20	144
		Tetrachloro-m-xyl	1	20	11.4	57		19	148
		Decachlorobiphen	2	20	8.12	41		20	144
		Tetrachloro-m-xyl	2	20	10.8	54		19	148
Q2600-05MS	STOCK-PILEMS	Decachlorobiphen	1	20	12.7	63		20	144
		Tetrachloro-m-xyl	1	20	12.6	63		19	148
		Decachlorobiphen	2	20	9.05	45		20	144
		Tetrachloro-m-xyl	2	20	11.6	58		19	148
Q2600-05MSD	STOCK-PILEMSD	Decachlorobiphen	1	20	12.5	62		20	144
		Tetrachloro-m-xyl	1	20	12.8	64		19	148
		Decachlorobiphen	2	20	9.09	45		20	144
		Tetrachloro-m-xyl	2	20	11.8	59		19	148
Q2600-09	END-OF-TRENCH	Decachlorobiphen	1	20	18.9	94		20	144
		Tetrachloro-m-xyl	1	20	8.80	44		19	148
		Decachlorobiphen	2	20	12.9	65		20	144
		Tetrachloro-m-xyl	2	20	9.81	49		19	148
I.BLK-PD089650.D	PIBLK-PD089650.D	Decachlorobiphen	1	20	16.1	80		57	171
		Tetrachloro-m-xyl	1	20	15.1	76		61	148
		Decachlorobiphen	2	20	13.1	66		57	171
		Tetrachloro-m-xyl	2	20	14.7	74		61	148