

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TACO01

Lab Code: ACE

SDG NO.: Q2600

Continuing Calib Date: 07/17/2025

Initial Calibration Date(s): 07/11/2025 07/11/2025

Continuing Calib Time: 10:04

Initial Calibration Time(s): 16:00 17:36

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.33	7.33	7.23	7.43	0.00
2,4-D	8.46	8.47	8.37	8.57	0.01
2,4,5-TP(Silvex)	9.35	9.35	9.25	9.45	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TACO01

Lab Code: ACE

SDG NO.: Q2600

Continuing Calib Date: 07/17/2025

Initial Calibration Date(s): 07/11/2025 07/11/2025

Continuing Calib Time: 10:04

Initial Calibration Time(s): 16:00 17:36

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.77	7.77	7.67	7.87	0.00
2,4-D	9.03	9.03	8.93	9.13	0.00
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00