

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** TACO01

**Lab Code:** ACE

**SDG NO.:** Q2600

**Continuing Calib Date:** 07/21/2025

**Initial Calibration Date(s):** 07/21/2025 07/21/2025

**Continuing Calib Time:** 17:51

**Initial Calibration Time(s):** 15:02 16:39

**GC Column:** RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.32	7.33	7.23	7.43	0.01
2,4-D	8.46	8.46	8.36	8.56	0.00
2,4,5-TP(Silvex)	9.34	9.34	9.24	9.44	0.00

**CALIBRATION VERIFICATION SUMMARY**

**Lab Name:** Alliance

**Contract:** TACO01

**Lab Code:** ACE

**SDG NO.:** Q2600

**Continuing Calib Date:** 07/21/2025

**Initial Calibration Date(s):** 07/21/2025 07/21/2025

**Continuing Calib Time:** 17:51

**Initial Calibration Time(s):** 15:02 16:39

**GC Column:** RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.77	7.77	7.67	7.87	0.00
2,4-D	9.02	9.02	8.92	9.12	0.00
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00