

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/17/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031100.D **Time Analyzed:** 17:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.348	9.253	9.453	741.040	712.500	4.0
2,4-D	8.462	8.367	8.567	777.450	705.000	10.3
2,4-DCAA	7.329	7.233	7.433	712.900	750.000	-4.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/17/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031100.D **Time Analyzed:** 17:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.932	9.832	10.032	609.110	712.500	-14.5
2,4-D	9.027	8.927	9.127	592.540	705.000	-16.0
2,4-DCAA	7.768	7.669	7.869	619.530	750.000	-17.4