

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031126.D **Time Analyzed:** 11:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.346	9.253	9.453	813.240	712.500	14.1
2,4-D	8.461	8.367	8.567	982.070	705.000	39.3
2,4-DCAA	7.327	7.233	7.433	639.190	750.000	-14.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/11/2025 07/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/18/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031126.D **Time Analyzed:** 11:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.926	9.832	10.032	675.120	712.500	-5.2
2,4-D	9.021	8.927	9.127	740.430	705.000	5.0
2,4-DCAA	7.763	7.669	7.869	658.150	750.000	-12.2