

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/21/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031164.D **Time Analyzed:** 17:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.343	9.242	9.442	702.680	712.500	-1.4
2,4-D	8.457	8.356	8.556	689.120	705.000	-2.3
2,4-DCAA	7.324	7.225	7.425	735.310	750.000	-2.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/21/2025 07/21/2025

Client Sample No.: CCAL06 **Date Analyzed:** 07/21/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031164.D **Time Analyzed:** 17:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.929	9.829	10.029	704.050	712.500	-1.2
2,4-D	9.024	8.924	9.124	685.120	705.000	-2.8
2,4-DCAA	7.766	7.666	7.866	730.430	750.000	-2.6