

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2600</u>	Analytical Method:	<u>8151A</u>
Client:	<u>T&A Construction Inc</u>	DataFile :	<u>PS031116.D</u>

		Sample				Rec		RPD		Limits		
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2592-02MSD	Client Sample ID:		WC-SOIL-20250711MSD								
	(Column 1)											
	2,4-D	50	0	70.1	ug/L	140	*	1		65	135	20
	2,4,5-TP(Silvex)	50	0	63.5	ug/L	127		1		62	139	20
Lab Sample ID:	Q2592-02MSD	Client Sample ID:		WC-SOIL-20250711MSD								
	(Column 2)											
	2,4-D	50	0	55.8	ug/L	112	*	1		65	135	20
	2,4,5-TP(Silvex)	50	0	52.6	ug/L	105		3		62	139	20