

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: TACO01

Lab Code: ACE

SDG NO.: Q2600

Instrument ID: ECD_L

Calibration Date(s): 07/07/2025 07/07/2025

Calibration Times: 10:53 11:49

GC Column: ZB-MR1

ID: 0.32 (mm)

LAB FILE ID:							
CF 100 = <u>PL096240.D</u>		CF 075 = <u>PL096241.D</u>					
CF 050 = <u>PL096242.D</u>		CF 025 = <u>PL096243.D</u>		CF 005 = <u>PL096244.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	2734760000	2660070000	2729730000	2840050000	3100690000	2813060000	6
Endrin	3278570000	3096260000	3152290000	3035270000	2988660000	3110210000	4
gamma-BHC (Lindane)	5277490000	4937170000	5003630000	4872450000	4807500000	4979650000	4
Heptachlor	4644220000	4480710000	4500180000	4479130000	4801850000	4581220000	3
Heptachlor epoxide	4361240000	4217620000	4223310000	4173120000	3770510000	4149160000	5
Methoxychlor	1520130000	1498290000	1535260000	1580890000	1742410000	1575400000	6
Tetrachloro-m-xylene	3687130000	3500570000	3606810000	3645250000	3805110000	3648970000	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: TACO01

Lab Code: ACE

SDG NO.: Q2600

Instrument ID: ECD_L

Calibration Date(s): 07/07/2025 07/07/2025

Calibration Times: 10:53 11:49

GC Column: ZB-MR2

ID: 0.32 (mm)

LAB FILE ID:		CF 100 =	<u>PL096240.D</u>	CF 075 =	<u>PL096241.D</u>		
CF 050 =		<u>PL096242.D</u>	CF 025 =	<u>PL096243.D</u>	CF 005 =	<u>PL096244.D</u>	
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
Decachlorobiphenyl	4787910000	4666610000	4872180000	5085100000	5520680000	4986500000	7
Endrin	6179120000	5963060000	6145890000	6105230000	6089210000	6096500000	1
gamma-BHC (Lindane)	8140460000	7776520000	7961280000	7854920000	7646170000	7875870000	2
Heptachlor	7661840000	7400290000	7666040000	7667210000	7643400000	7607760000	2
Heptachlor epoxide	6588350000	6451450000	6691850000	6724070000	6968780000	6684900000	3
Methoxychlor	2949250000	2893130000	3027760000	3113420000	3205310000	3037770000	4
Tetrachloro-m-xylene	5771200000	5549810000	5746090000	5799260000	5989730000	5771220000	3