

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TACO01

Lab Code: ACE

SDG NO.: Q2600

Continuing Calib Date: 07/18/2025

Initial Calibration Date(s): 07/07/2025

07/07/2025

Continuing Calib Time: 09:07

Initial Calibration Time(s): 10:53

11:49

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 9.02       | 9.02      | 8.92                   | 9.12 | 0.00       |
| Tetrachloro-m-xylene | 3.54       | 3.54      | 3.44                   | 3.64 | 0.00       |
| gamma-BHC (Lindane)  | 4.31       | 4.31      | 4.21                   | 4.41 | 0.00       |
| Heptachlor           | 4.91       | 4.91      | 4.81                   | 5.01 | 0.00       |
| Heptachlor epoxide   | 5.66       | 5.67      | 5.57                   | 5.77 | 0.01       |
| Endrin               | 6.55       | 6.55      | 6.45                   | 6.65 | 0.00       |
| Methoxychlor         | 7.46       | 7.46      | 7.36                   | 7.56 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: TACO01

Lab Code: ACE

SDG NO.: Q2600

Continuing Calib Date: 07/18/2025

Initial Calibration Date(s): 07/07/2025 07/07/2025

Continuing Calib Time: 09:07

Initial Calibration Time(s): 10:53 11:49

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|------------|-----------|------------------------|------|------------|
| Decachlorobiphenyl   | 7.99       | 8.00      | 7.90                   | 8.10 | 0.01       |
| Tetrachloro-m-xylene | 2.83       | 2.83      | 2.73                   | 2.93 | 0.00       |
| gamma-BHC (Lindane)  | 3.67       | 3.67      | 3.57                   | 3.77 | 0.00       |
| Heptachlor           | 4.02       | 4.02      | 3.92                   | 4.12 | 0.01       |
| Heptachlor epoxide   | 4.80       | 4.80      | 4.70                   | 4.90 | 0.00       |
| Endrin               | 5.71       | 5.71      | 5.61                   | 5.81 | 0.00       |
| Methoxychlor         | 6.68       | 6.68      | 6.58                   | 6.78 | 0.00       |