

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096443.D **Time Analyzed:** 09:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.017	8.920	9.120	47.180	50.000	-5.6
Endrin	6.545	6.445	6.645	55.790	50.000	11.6
gamma-BHC (Lindane)	4.313	4.213	4.413	54.690	50.000	9.4
Heptachlor	4.905	4.806	5.006	53.350	50.000	6.7
Heptachlor epoxide	5.664	5.565	5.765	55.500	50.000	11.0
Methoxychlor	7.464	7.364	7.564	51.360	50.000	2.7
Tetrachloro-m-xylene	3.536	3.438	3.638	53.670	50.000	7.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096443.D **Time Analyzed:** 09:07

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.993	7.896	8.096	48.860	50.000	-2.3
Endrin	5.710	5.612	5.812	54.570	50.000	9.1
gamma-BHC (Lindane)	3.666	3.567	3.767	57.380	50.000	14.8
Heptachlor	4.015	3.917	4.117	56.880	50.000	13.8
Heptachlor epoxide	4.800	4.702	4.902	55.240	50.000	10.5
Methoxychlor	6.682	6.584	6.784	51.220	50.000	2.4
Tetrachloro-m-xylene	2.828	2.729	2.929	56.320	50.000	12.6