

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096453.D **Time Analyzed:** 11:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.018	8.920	9.120	46.650	50.000	-6.7
Endrin	6.544	6.445	6.645	54.890	50.000	9.8
gamma-BHC (Lindane)	4.312	4.213	4.413	53.830	50.000	7.7
Heptachlor	4.904	4.806	5.006	53.120	50.000	6.2
Heptachlor epoxide	5.663	5.565	5.765	54.380	50.000	8.8
Methoxychlor	7.463	7.364	7.564	49.480	50.000	-1.0
Tetrachloro-m-xylene	3.536	3.438	3.638	52.660	50.000	5.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: TACO01
Lab Code: ACE
SDG NO.: Q2600
GC Column: ZB-MR2
ID: 0.32 (mm)
 Initi. Calib. Date(s): 07/07/2025 07/07/2025

Client Sample No.: CCAL02
Date Analyzed: 07/18/2025
Lab Sample No.: PSTDCCC050
Data File : PL096453.D
Time Analyzed: 11:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.992	7.896	8.096	48.810	50.000	-2.4
Endrin	5.709	5.612	5.812	55.610	50.000	11.2
gamma-BHC (Lindane)	3.665	3.567	3.767	56.110	50.000	12.2
Heptachlor	4.015	3.917	4.117	55.540	50.000	11.1
Heptachlor epoxide	4.799	4.702	4.902	54.370	50.000	8.7
Methoxychlor	6.681	6.584	6.784	50.050	50.000	0.1
Tetrachloro-m-xylene	2.828	2.729	2.929	55.020	50.000	10.0