

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096465.D **Time Analyzed:** 14:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.018	8.920	9.120	46.510	50.000	-7.0
Endrin	6.544	6.445	6.645	54.470	50.000	8.9
gamma-BHC (Lindane)	4.311	4.213	4.413	53.800	50.000	7.6
Heptachlor	4.904	4.806	5.006	53.460	50.000	6.9
Heptachlor epoxide	5.663	5.565	5.765	55.420	50.000	10.8
Methoxychlor	7.463	7.364	7.564	49.330	50.000	-1.3
Tetrachloro-m-xylene	3.536	3.438	3.638	52.910	50.000	5.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** TACO01
Lab Code: ACE **SDG NO.:** Q2600
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/18/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL096465.D **Time Analyzed:** 14:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.993	7.896	8.096	47.190	50.000	-5.6
Endrin	5.709	5.612	5.812	55.030	50.000	10.1
gamma-BHC (Lindane)	3.665	3.567	3.767	55.780	50.000	11.6
Heptachlor	4.014	3.917	4.117	55.060	50.000	10.1
Heptachlor epoxide	4.799	4.702	4.902	53.660	50.000	7.3
Methoxychlor	6.681	6.584	6.784	48.710	50.000	-2.6
Tetrachloro-m-xylene	2.827	2.729	2.929	54.910	50.000	9.8