

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
 Method File : 8270E-BP070325.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jul 03 16:05:44 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BP025071.D 5 =BP025072.D 10 =BP025073.D 20 =BP025074.D 40 =BP025075.D 50 =BP025076.D 60 =BP025077.D 80 =BP025078.D

	Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.470	0.452	0.457	0.435	0.420	0.428	0.409	0.439	4.99	
3)	Pyridine	1.176	1.153	1.243	1.228	1.171	1.239	1.203	1.202	3.00	
4)	n-Nitrosodimet...	0.500	0.531	0.506	0.493	0.522	0.500	0.509	0.509	2.85	
5) S	2-Fluorophenol	1.070	1.098	1.179	1.134	1.127	1.168	1.127	1.129	3.33	
6)	Aniline	1.334	1.324	1.430	1.388	1.371	1.428	1.359	1.376	3.05	
7) S	Phenol-d6	1.428	1.407	1.520	1.458	1.472	1.543	1.482	1.473	3.26	
8)	2-Chlorophenol	1.254	1.207	1.320	1.256	1.250	1.294	1.259	1.263	2.83	
9)	Benzaldehyde	0.903	0.894	0.724	0.684	0.641	0.769	15.84			
10) C	Phenol	1.473	1.461	1.551	1.472	1.485	1.540	1.493	1.497	2.36	
11)	bis(2-Chloroet...	1.187	1.139	1.200	1.122	1.129	1.172	1.115	1.152	2.95	
12)	1,3-Dichlorobe...	1.461	1.434	1.485	1.379	1.366	1.413	1.356	1.413	3.48	
13) C	1,4-Dichlorobe...	1.500	1.432	1.505	1.399	1.384	1.424	1.374	1.431	3.70	
14)	1,2-Dichlorobe...	1.454	1.388	1.445	1.349	1.329	1.368	1.303	1.377	4.12	
15)	Benzyl Alcohol	0.887	0.991	0.957	0.990	1.058	1.015	0.983	5.86		
16)	2,2'-oxybis(1-...	1.587	1.520	1.570	1.433	1.462	1.497	1.415	1.498	4.40	
17)	2-Methylphenol	0.887	0.908	1.007	0.969	0.974	1.021	0.983	0.964	5.11	
18)	Hexachloroethane	0.476	0.441	0.480	0.460	0.463	0.481	0.462	0.466	3.04	
19) P	n-Nitroso-di-n...	0.815	0.862	0.854	0.938	0.865	0.871	0.912	0.872	0.874	4.26
20)	3+4-Methylphenols	1.246	1.383	1.334	1.341	1.428	1.370	1.350	4.55		
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.480	0.481	0.502	0.465	0.456	0.480	0.447	0.473	3.88	
23) S	Nitrobenzene-d5	0.267	0.277	0.314	0.307	0.308	0.323	0.308	0.301	6.84	
24)	Nitrobenzene	0.280	0.293	0.319	0.310	0.307	0.327	0.308	0.306	5.13	
25)	Isophorone	0.525	0.534	0.587	0.558	0.571	0.597	0.561	0.562	4.67	
26) C	2-Nitrophenol	0.086	0.093	0.117	0.133	0.146	0.160	0.159	0.128	23.51	
27)	2,4-Dimethylph...	0.199	0.195	0.218	0.208	0.213	0.224	0.210	0.210	4.79	
28)	bis(2-Chloroet...	0.378	0.371	0.401	0.374	0.379	0.397	0.371	0.382	3.25	
29) C	2,4-Dichloroph...	0.242	0.263	0.298	0.299	0.304	0.321	0.306	0.291	9.48	
30)	1,2,4-Trichlor...	0.319	0.319	0.331	0.319	0.316	0.335	0.314	0.322	2.53	
31)	Naphthalene	1.057	1.030	1.081	1.011	1.005	1.049	0.975	1.030	3.47	
32)	Benzoic acid	0.115	0.152	0.186	0.199	0.225	0.231	0.185	24.07		
33)	4-Chloroaniline	0.358	0.370	0.399	0.390	0.381	0.411	0.381	0.384	4.58	
34) C	Hexachlorobuta...	0.187	0.180	0.188	0.182	0.184	0.191	0.180	0.185	2.31	
35)	Caprolactam	0.087	0.101	0.101	0.098	0.108	0.102	0.099	7.23		
36) C	4-Chloro-3-met...	0.259	0.279	0.321	0.309	0.309	0.332	0.316	0.304	8.35	
37)	2-Methylnaphth...	0.724	0.711	0.766	0.718	0.715	0.747	0.706	0.727	2.99	
38)	1-Methylnaphth...	0.707	0.704	0.753	0.692	0.691	0.736	0.678	0.709	3.74	

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
 Method File : 8270E-BP070325.M


39) I	Acenaphthene-d10	-----ISTD-----								
40)	1,2,4,5-Tetrac...	0.554	0.549	0.584	0.580	0.578	0.610	0.568	0.575	3.56
41) P	Hexachlorocycl...	0.120	0.135	0.149	0.162	0.171	0.164	0.150		13.13
42) S	2,4,6-Tribromo...	0.194	0.214	0.240	0.246	0.244	0.266	0.249	0.236	10.22
43) C	2,4,6-Trichlor...	0.271	0.295	0.355	0.370	0.368	0.399	0.379	0.348	13.51
44)	2,4,5-Trichlor...	0.326	0.349	0.406	0.411	0.412	0.441	0.416	0.394	10.41
45) S	2-Fluorobiphenyl	1.338	1.310	1.361	1.279	1.260	1.311	1.177	1.291	4.70
46)	1,1'-Biphenyl	1.482	1.436	1.510	1.454	1.435	1.490	1.380	1.455	2.99
47)	2-Chloronaphth...	1.097	1.089	1.141	1.091	1.069	1.121	1.048	1.094	2.83
48)	2-Nitroaniline	0.157	0.185	0.235	0.259	0.259	0.289	0.273	0.237	20.49
49)	Acenaphthylene	1.582	1.622	1.771	1.701	1.660	1.761	1.620	1.674	4.36
50)	Dimethylphthalate	1.364	1.376	1.469	1.408	1.365	1.482	1.337	1.400	3.99
51)	2,6-Dinitrotol...	0.199	0.238	0.281	0.295	0.291	0.313	0.294	0.273	14.58
52) C	Acenaphthene	1.085	1.074	1.142	1.066	1.047	1.111	1.027	1.079	3.59
53)	3-Nitroaniline	0.204	0.248	0.297	0.321	0.310	0.345	0.323	0.293	16.81
54) P	2,4-Dinitrophenol	0.077	0.098	0.117	0.120	0.137	0.137	0.114		20.49
55)	Dibenzofuran	1.829	1.814	1.866	1.768	1.720	1.813	1.655	1.781	4.07
56) P	4-Nitrophenol	0.212	0.260	0.285	0.274	0.301	0.289	0.270		11.73
57)	2,4-Dinitrotol...	0.231	0.285	0.364	0.392	0.387	0.430	0.406	0.356	20.13
58)	Fluorene	1.382	1.384	1.467	1.393	1.335	1.403	1.278	1.377	4.27
59)	2,3,4,6-Tetrac...	0.305	0.318	0.370	0.370	0.363	0.397	0.377	0.357	9.32
60)	Diethylphthalate	1.368	1.376	1.468	1.429	1.366	1.477	1.358	1.406	3.64
61)	4-Chlorophenyl...	0.711	0.687	0.710	0.679	0.659	0.700	0.626	0.682	4.47
62)	4-Nitroaniline	0.221	0.267	0.321	0.342	0.319	0.354	0.324	0.307	15.26
63)	Azobenzene	1.251	1.275	1.328	1.265	1.211	1.265	1.171	1.252	3.98
64) I	Phenanthrene-d10	-----ISTD-----								
65)	4,6-Dinitro-2-...	0.054	0.069	0.081	0.089	0.097	0.098	0.082		21.25
66) c	n-Nitrosodiphe...	0.557	0.548	0.597	0.573	0.574	0.592	0.565	0.572	3.11
67)	4-Bromophenyl-...	0.191	0.190	0.213	0.205	0.208	0.214	0.213	0.205	5.14
68)	Hexachlorobenzene	0.255	0.240	0.255	0.247	0.248	0.257	0.249	0.250	2.28
69)	Atrazine	0.155	0.167	0.183	0.172	0.171	0.172	0.154	0.168	6.10
70) C	Pentachlorophenol	0.131	0.163	0.172	0.178	0.186	0.184	0.169		12.03
71)	Phenanthrene	1.102	1.084	1.104	1.068	1.030	1.082	1.011	1.069	3.34
72)	Anthracene	0.998	0.996	1.092	1.036	1.021	1.062	0.971	1.025	4.10
73)	Carbazole	0.954	0.994	1.057	1.029	0.990	1.017	0.976	1.002	3.45
74)	Di-n-butylphth...	0.929	0.994	1.147	1.187	1.177	1.228	1.152	1.116	9.89
75) C	Fluoranthene	1.170	1.222	1.330	1.290	1.249	1.268	1.215	1.249	4.22
76) I	Chrysene-d12	-----ISTD-----								
77)	Benzidine	0.177	0.354	0.571	0.420	0.288	0.310	0.353		37.82
78)	Pyrene	1.177	1.184	1.288	1.231	1.251	1.302	1.238	1.239	3.81
79) S	Terphenyl-d14	0.943	0.919	0.978	0.914	0.915	0.936	0.764	0.910	7.50
80)	Butylbenzylpht...	0.228	0.276	0.377	0.440	0.471	0.519	0.499	0.402	28.01
81)	Benzo(a)anthra...	1.146	1.175	1.293	1.217	1.207	1.258	1.189	1.212	4.11
82)	3,3'-Dichlorob...	0.249	0.340	0.390	0.399	0.414	0.397	0.365		16.97
83)	Chrysene	1.169	1.147	1.223	1.159	1.157	1.191	1.118	1.166	2.88
84)	Bis(2-ethylhex...	0.460	0.520	0.646	0.703	0.737	0.764	0.720	0.650	17.89
85) c	Di-n-octyl pht...	0.609	0.876	1.094	1.206	1.292	1.248	1.054		25.04

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP070325.M

86) I	Perylene-d12	-----ISTD-----									
87)	Indeno(1,2,3-c...	1.003	1.101	1.318	1.331	1.366	1.433	1.366	1.274	12.44	
88)	Benzo(b)fluora...	1.029	1.063	1.215	1.201	1.188	1.266	1.187	1.164	7.36	
89)	Benzo(k)fluora...	1.070	1.129	1.249	1.153	1.157	1.218	1.133	1.158	5.11	
90) C	Benzo(a)pyrene	0.817	0.868	1.035	1.033	1.035	1.108	1.040	0.991	10.70	
91)	Dibenzo(a,h)an...	0.840	0.921	1.098	1.097	1.126	1.172	1.118	1.053	11.66	
92)	Benzo(g,h,i)pe...	0.915	0.967	1.126	1.127	1.142	1.197	1.142	1.088	9.58	

(#) = Out of Range


2-Nitrophenol



Response = $1.596 \times 10^{-1} \times \text{Amt} - 2.635 \times 10^{-2}$
 Coef of Det (r^2) = 0.992912 Curve Fit: wlr(1/a)
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M
 Calibration Table Last Updated: Thu Jul 03 16:05:44 2025

Benzoic acid

Response Ratio



$$\text{Response} = 2.405\text{e-}001 * \text{Amt} - 7.461\text{e-}002$$


Coef of Det (r^2) = 0.993174 Curve Fit: wlr(1/a)

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M

Calibration Table Last Updated: Thu Jul 03 16:05:44 2025

2-Nitroaniline

Response Ratio



$$\text{Response} = 2.843\text{e-}001 * \text{Amt} - 3.930\text{e-}002$$


Coef of Det (r^2) = 0.996989 Curve Fit: wlr(1/a)

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M

Calibration Table Last Updated: Thu Jul 03 16:05:44 2025

2,4-Dinitrotoluene

Response Ratio




$$\text{Response} = 4.240\text{e-}001 * \text{Amt} - 5.591\text{e-}002$$

Coef of Det (r^2) = 0.997730 Curve Fit: wlr(1/a)

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M


Calibration Table Last Updated: Thu Jul 03 16:05:44 2025

2,4-Dinitrophenol



Response = $1.431e-001 * \text{Amt} - 3.852e-002$
 Coef of Det (r^2) = 0.995026 Curve Fit: wlr(1/a)
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M
 Calibration Table Last Updated: Thu Jul 03 16:05:44 2025


4,6-Dinitro-2-methylphenol



Response = 1.030e-001 * Amt - 2.873e-002
 Coef of Det (r^2) = 0.995315 Curve Fit: wlr(1/a)
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M
 Calibration Table Last Updated: Thu Jul 03 16:05:44 2025

Butylbenzylphthalate

Response Ratio



$$\text{Response} = 5.169\text{e-}001 * \text{Amt} - 9.516\text{e-}002$$


Coef of Det (r^2) = 0.994647 Curve Fit: wlr(1/a)

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M

Calibration Table Last Updated: Thu Jul 03 16:05:44 2025

Di-n-octyl phthalate

Response Ratio




$$\text{Response} = 1.366\text{e}+000 * \text{Amt} - 4.164\text{e}-001$$

Coef of Det (r^2) = 0.996792 Curve Fit: wlr(1/a)

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M

Calibration Table Last Updated: Thu Jul 03 16:05:44 2025

2,4-Dinitrophenol



Response = $1.431e-001 * \text{Amt} - 3.852e-002$
 Coef of Det (r^2) = 0.995026 Curve Fit: wlr(1/a)
 Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP070325.M
 Calibration Table Last Updated: Thu Jul 03 16:05:44 2025