

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2600

Client ID : SSTDCCC040

Date Analyzed: 07/16/2025

Lab File ID: BP025155.D

Time Analyzed: 12:42

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	564997	7.443	2237580	10.20	1465070	14.10
UPPER LIMIT	1129990	7.943	4475160	10.696	2930140	14.595
LOWER LIMIT	282499	6.943	1118790	9.696	732535	13.595
EPA SAMPLE NO.						
01 TRENCH	670552	7.44	2634040	10.19	1646660	14.10
02 STOCK-PILE	625573	7.44	2410870	10.19	1564530	14.10
03 END-OF-TRENCH	582317	7.45	2493690	10.20	1630810	14.10

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.